Technical Datasheet

Date: 18.02.20

Page 1 of 1



		EN ISO 13485 certified
Article - No:	04-032-0400	Revision: 4
Remarks	3-linkages	single use only
	Specifica	tion
Product PCR microcentrifuge tube PP, 0,2ml, 8-strips, individually attached flat & transpa		
Tioduct	linkages, highly transparent, np pcr ready, max. RCF 20.000g, autocl. up to 121°C, CE/IV	
Classification	CE/IVD	
Colour	highly transparent	
Nominal Capacity	0,2ml	
Material Characteristics	polypropylene (PP)	
Temperature range	-80°C - 121°C	
Autoclavability	autocl. up to 121°C	
Centrifugation	max. RCF 20.000g	
Product Characteristics	biologically inert	
	physiologically harmless	
	no heavy metals according to EC/94/62	
	Food contact: raw material in compliance with regulation EC 1935/2004	
Purity Grade	free of PCR inhibitors	
	free of human DNA und RNA	
	free of DNase and RNase	
	free of pyrogens	
	free of endotoxins	
Shelf - Life	not limited at controlled storage conditions	
Packaging	pcs per inner cs: 120	
	pcs per carton: 1.200	
	pcs per pallet: 57.600	
Storage	dry condition: 40%-60% rel. humidity	
	bags / cases orig. closed	
	storage temp. +2°C to max. +30°C	
	pal. not stackable	
Approved by nerbe plus	quality department	

nerbe plus GmbH & Co. KG Porschestr. 25 info@nerbe-plus.de D-21423 Winsen/Luhe www.nerbe-plus.de Tel: +49 (0)4171 7849 0 HRA Lüneburg 203240 Fax:+49 (0)4171 6 45 45 USt.-ID: DE 32583 0652

Pers, haftende Gesellschafterin: NERBE LAB GmbH, Winsen/Luhe GF: Heike Nerbe HRB Lüneburg 207694 Deutsche Bank AG Kto.0204 41200 IBAN: DE41 2407 0075 0020 4412 00 SWIFT / BIC DEUTDE2H240

Hamburger Sparkasse Kto, 1387 140831 IBAN: DE56 2005 0550 1387 1408 31 SWIFT / BIC HASPDEHHXXX

