

Technical Datasheet

Date: 18.02.20

Page 1 of 2

EN ISO 13485 certified		
Article - No:	04-032-0300	Revision: 5
Remarks	breakable strip	single use only

Specification

Product	PCR microcentrifuge tube PP, 0,2ml, 8-strips, individually attached flat & transparent caps, breakable strip, highly transparent, np pcr ready max. RCF 20.000g, autocl. up to 121°C, CE/IVD
Classification	CE/IVD
Colour	highly transparent
Nominal Capacity	0,2ml
Material Characteristics	polypropylene (PP)
Temperature range	-80°C - 121°C
Autoclavability	autocl. up to 121°C
Centrifugation	max. RCF 20.000g
Product Characteristics	biologically inert
	physiologically harmless
	no heavy metals according to EC/94/62
	Food contact: raw material in compliance with regulation EC 1935/2004
Purity Grade	free of PCR inhibitors
	free of human DNA und RNA
	free of DNase and RNase
	free of pyrogens
	free of endotoxins
Shelf - Life	not limited at controlled storage conditions
Packaging	pcs per inner cs: 120
	pcs per carton: 1.200
	pcs per pallet: 57.600
Storage	dry condition: 40%-60% rel. humidity
	bags / cases orig. closed
	storage temp. +2°C to max. +30°C
	pal. not stackable

Technical Datasheet

Date: 18.02.20

Page 2 of 2



EN ISO 13485 certified

Approved by nerbe plus quality department

00-000-0001/02.2020

nerbe plus GmbH & Co. KG
Porschestr. 25
D-21423 Winsen/Luhe
Tel: +49 (0)4171 7849 0
Fax: +49 (0)4171 6 45 45

info@nerbe-plus.de
www.nerbe-plus.de
HRA Lüneburg 203240
USt.-ID: DE 32583 0652

Pers. haftende Gesellschafterin:
NERBE LAB GmbH, Winsen/Luhe
GF: Heike Nerbe
HRB Lüneburg 207694

Deutsche Bank AG
Kto. 0204 41200
IBAN: DE41 2407 0075 0020 4412 00
SWIFT / BIC DEUTDE2H240

Hamburger Sparkasse
Kto. 1387 140831
IBAN: DE56 2005 0550 1387 1408 31
SWIFT / BIC HASPDEHHXXX

