

# Technical Datasheet

Date: 18.02.20

Page 1 of 1



EN ISO 13485 certified

EN ISO 13485 certified		
Article - No:	04-212-3000	Revision: 5
Remarks	piercable cap	single use only

## Specification

Product	Microcentrifuge tube PP, 1,5ml, attached cap, heat-resistant, transparent, graduated, with writing area, np pcr ready, max. RCF 20.000g, autoclavable up to 121°C, CE/IVD
Classification	CE/IVD
Colour	transparent
Graduation	grad., writ. area
Nominal Capacity	1,5ml
Material Characteristics	polypropylene (PP)
Temperature range	-80°C - 121°C
Autoclavability	autocl. up to 121°C
Centrifugation	max. RCF 20.000g
Product Characteristics	biologically inert
	physiologically harmless
	constantly usable at temperature of 80 °C up to 100 °C
Purity Grade	free of PCR inhibitors
	free of human DNA und RNA
	free of DNase and RNase
	free of pyrogens
	free of endotoxins
Packaging	pcs per bag: 200
	pcs per carton: 5.000
	pcs per pallet: 100.000
Storage	dry condition: 40%-60% rel. humidity
	bags / cases orig. closed
	storage temp. +2°C to max. +30°C
	pal. not stackable
Approved by nerbe plus quality department	

00-000-0001/02.2020

nerbe plus GmbH & Co. KG  
Porschestra. 25  
D-21423 Winsen/Luhe  
Tel: +49 (0)4171 7849 0  
Fax: +49 (0)4171 6 45 45

info@nerbe-plus.de  
www.nerbe-plus.de  
HRA Lüneburg 203240  
USt.-ID: DE 32583 0652

Pers. haftende Gesellschafterin:  
NERBE LAB GmbH, Winsen/Luhe  
GF: Heike Nerbe  
HRB Lüneburg 207694

Deutsche Bank AG  
Kto. 0204 41200  
IBAN: DE41 2407 0075 0020 4412 00  
SWIFT / BIC DEUTDE2H240

Hamburger Sparkasse  
Kto. 1387 140831  
IBAN: DE56 2005 0550 1387 1408 31  
SWIFT / BIC HASPDEHHXXX

